

CERTIFICATE OF ANALYSIS

Product: qPCR 2x Master Mix

Catalog No: P501, P502, P503, P503xl

Lot No: P501122023

Date of Expiry: 12/2023

Composition: qPCR 2x Master Mix contains: 20 mM Tris-HCl, pH 8.8 (at 25°C), 100 mM KCl, 3 mM MgCl₂, 400 μM dATP, 400 μM dCTP, 400 μM dGTP, 400 μM dTTP, Taq DNA polymerase (50 U/ml), monoclonal antibody anti-Taq (38 mM), stabilizers and additives.

Supplied with: PCR Ultra H₂O (Cat. No. P040).

Storage temperature: For short terms (days) at 4°C ± 3°C.
For long terms at -20 ± 5°C. Material can be repeatedly defrosted.

Functional Test: The lot has been tested for the ability to amplify a fragment of genomic DNA using the following tconditions:

Test conditions:

Volume*	Reagent	Final concentration
12.5 μl	qPCR 2x Master Mix	1x qPCR 2x Master Mix
0.5 μl	Forward primer	50 μM 5' primer, 5'-ATGAACCCAGCCATCAGCG-3'
0.5 μl	Reverse primer	50 μM 3' primer 5'-GGGTAAGGACCTTGATATAGG-3'
1 μl	Template DNA	containing 80 ng of mouse genomic DNA
10.5 μl	PCR Ultra H ₂ O	(to a final volume 25 μl)

Cycling conditions on LightCycler 96, ROCHE:

	Temperature	Time	Number of cycles
Initial denaturation	94°C	3 min	1
Denaturation	94°C	15 s	35
Annealing of primers	55°C	15 s	
Extension	72°C	1 min	
Final extension	72°C	7 min	1
Cooling	22°C		

Result: As expected, As expected, electrophoresis of the PCR product on agarose gel revealed one band of 864 bp.

FOR RESEARCH USE**APPROVED DATE:** 04.08.2021

Manager: Hana Těšitelová